

1320 Research Park Drive  
Manhattan, Kansas 66502  
(785) 564-6700



900 SW Jackson, Room 456  
Topeka, Kansas 66612  
(785) 296-3556

Jackie McClaskey, Secretary

Governor Sam Brownback

July 25, 2016

Terry Kluge, Senior Regulatory Compliance Engineer  
Badger Meter  
4545 W Brown Deer Rd  
PO Box 245036  
Milwaukee, WI 53224-9536

Dear Mr. Kluge,

I am pleased to inform you that the information included with your revised request, received in this office on July 12, 2016, along with additional information received through several emails, for the following water flowmeter model series, has been reviewed and it has been determined those water flowmeter model series meet the specifications of the Chief Engineer as adopted in K.A.R. 5-1-4, subject to the conditions listed below, and will be included on the List of Certified Water Flowmeters maintained in accordance with K.A.R. 5-1-12.

**M1000 Eletromagnetic Flow Meter (1/4" through 8")**  
**M2000 Eletromagnetic Flow Meter (1/4" through 54")**  
**M5000 Eletromagnetic Flow Meter (1/2" through 24")**

The meters will be shipped with a seal in place with only an administrative level password which the user will not have access to. There will be no user or service level password. If additional programming is needed after shipping, a Badger sales representative will do the additional setup. After the setup is accomplished, the representative will lock the meter with the administrative password and attach a seal screw kit (lead seal with seal wire) to complete the installation. The seal kit will be easily identified as a Badger Meter product. The administrative password will be retained only by Badger Meter and will in no case be revealed to the user.

Placement on the list means the above named model series, when installed properly and measuring accurately, will be accepted to meet water flowmeter requirements imposed by the Chief Engineer.

Sincerely yours,

David W. Barfield, P.E.  
Chief Engineer

Pc: Cindy Kransler, Badger Meter, 4545 W. Brown Deer Run, Milwaukee, WI 53244  
John Huff, Badger Meter, 4545 W. Brown Deer Run, Milwaukee, WI 53244  
Topeka Field Office  
Stafford Field Office  
Stockton Field Office  
Garden City Field Office